



UK Declaration of Conformity

Technical File No.

003

Document No. Q1182


Revision. 1.0

For Medical Device / Device Family (Generic Device Group)	Alert-iT Pager
Legal Manufacturer; Name and Address	iTs Designs Ltd., (Trading as: Alert-IT Care Alarms) Fernie House (Unit 3) Coalville Business Park Coalville, Leicestershire LE67 3NR UK
Trade Name or Registered Trade Mark	P202 - Alert-iT Pager
Product Codes & Labelling Descriptions	As per ANNEX A
Device Classification; as per Annex IX of UK MDR 2002	Class 1 Non-sterile, Non-measuring, Non-reusable surgical instrument, Non- custom made.  Rule 12
Intended Use	The Alert-iT Pager is intended to assist in caregiving by notifying the carer to attend the patient in order the verify their wellbeing
Harmonised Standards, or Common Specifications Applied	As per ANNEX A
Medical Device Regulation Conformity Assessment Procedure Undertaken	Manufacturer self-certification through issuance of a UK Declaration of Conformity in accordance with UK MDR 2002 (SI 2002 no.618, amended), Annex VII.

	EU Declaration of Conformity	
	Technical File No.	003
	Document No. Q1182	Revision. 1.0

## Statement of Undertaking:

The undersigned declares that the UK Declaration of Conformity is issued under the sole responsibility of the manufacturer, and that the device(s) listed within are in conformity with UK MDR 2002 (SI 2002 no.618, amended)


	Name	Title / Function	Location	Date	Signature
<b>Signed by:</b>	Rick Gunn	Operations Director	iTs Designs Ltd., (Trading as: Alert-IT Care Alarms)	30 <sup>th</sup> April 2025	
<b>On behalf of</b>	iTs Designs Ltd., (Trading as: Alert-IT Care Alarms) Ferne House (Unit 3) Coalville Business Park Coalville Leicestershire LE67 3NR UK				

## Annex A – Associated Model Identities

Product Code	Labelling Description
P202AAA	Standard 4 Channel MD Pager (434 MHz)
P202BAA	Advanced 128 Channel MD Pager (434 MHz)
P202ABA	Standard 4 Channel MD Pager (869 MHz)
P202BBA	Advanced 128 Channel MD Pager (869 MHz)

## Annex B –Standards or Common Specifications Applied

Harmonised Standard or Common Specification	Description
EN ISO 14971:2019	Application of Risk management to medical devices.
EN ISO 13485:2016	Medical devices. Quality management systems. Requirements for regulatory purposes
EN 62304:2006	Medical device software - Software life-cycle processes

	UK Declaration of Conformity	
	Technical File No.	003
	Document No. Q1182	Revision. 1.0

EN ISO 62366-1:2015	Medical devices. Application of usability engineering to medical devices
EN 60601-1:2006/A2:2021	Medical electrical equipment - Part 1: General requirements for basic safety and essential performance
EN 60601-1-2:2015+A1: 2021	Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral Standard: Electromagnetic disturbances - Requirements and tests
EN 60601-1-8:2007+A2:2021	Medical electrical equipment - Part 1-8: General requirements for basic safety and essential performance - Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems
EN 60601-1-11:2015+A1:2021	Medical electrical equipment - Part 1-11: General requirements for basic safety and essential performance - Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment.
EN 62133-1:2017	Secondary cells and batteries containing alkaline or other non-acid electrolytes - Safety requirements for portable sealed secondary cells, and for batteries made from them, for use in portable applications - Part 1: Nickel systems
EN ISO 15223-1:2021	Medical devices. Symbols to be used with medical device labels, labelling and information to be supplied. General requirements
EN ISO 20417:2021	Medical devices — Information to be supplied by the manufacturer

**END OF DOCUMENT**