

EC Declaration of Conformity




Manufacturer:

iTs Designs Ltd
 Atherstone House
 Merry Lees Industrial Estate
 Desford
 Leicestershire LE3 3LR
 England

Issue Date	21/8/2009	updated 20/12/16
Device(s)	P163ACB P163BAB	Wired Touch-iT with Switch Input Wired Sense-iT Bed Vacation
Accessories	P160 P143C P143G P144C P149A,B P171 P159B P145 Miscellaneous switch devices	NiMH Battery, Size AA, 1.25v Bed Mat Occupancy Sensor Bed Ribbon Occupancy Sensor Bed Leg Occupancy Sensor Chair Occupancy Sensor Range of power supplies Bed/Wall Mounting Bracket Range of Nurse Call Leads Passive switching devices covered by EN12182
Description	The P163 models cover a range of small monitors that detect abnormal conditions suggesting the user is at risk. These cover bed vacation, chair vacation, panic. The alarms are transmitted by Safelink Radio or to a connected Nurse Call or Telecare system	
Directives	2014/30/EU 2014/35/EU 2011/65/EU	EMC Directive Low Voltage Permitted Materials (UK)
Classification	Safety Class II	
Standards Applied	EN 14971:2012 EN 60950-1:2006+A2:2013 EN 60601-1-2:2015 EN 12182:2012	Risk Assessment Safety Medical Device EMC Assistive Technology

Declaration:
 We certify that the apparatus to which this declaration relates conforms with the relevant requirements set out in the above referenced EC Directive(s) and standards

Signatory:



D Godfrey BSc(Hons) FIEE Ceng