

EC Declaration of Conformity




Manufacturer:

iTs Designs Ltd
 Atherstone House
 Merry Lees Industrial Estate
 Desford
 Leicestershire LE3 3LR
 England

Issue Date	21/8/2009	updated 19/8/16
Device(s)	P163ABB	Wired Enuresis Alarm
Accessories	P160 P141E P142A P142 P113 P159B P145	NiMH Battery, Size AA, 1.25v Connecting Cable Instant Moisture Sensor Absorbent Moisture Sensor High protection power supplies Bed/Wall Mounting Bracket Range of Nurse Call Leads
Description	The P163 models cover a range of small monitors that detect moisture on a sensing sheet suggesting the user is at risk. The alarms are transmitted by Safelink Radio or to a connected Nurse Call or Telecare system	
Directives	2014/30/EU 2014/35/EU 2011/65/EU	EMC Directive Low Voltage Permitted Materials (UK)
Classification	Safety Class II	
Standards Applied	EN 14971:2012 EN 60950-1:2006+A2:2013 EN 60601-1-2:2015 EN 12182:2012	Risk Assessment Safety Medical Device EMC Assistive Technology

Declaration:
 We certify that the apparatus to which this declaration relates conforms with the relevant requirements set out in the above referenced EC Directive(s) and standards

Signatory:



D Godfrey BSc(Hons) FIEE Ceng
 Managing Director