

EC Declaration of Conformity



Manufacturer:

iTs Designs Ltd
Atherstone House
Merry Lees Industrial Estate
Desford
Leicestershire LE3 3LR
England

Issue Date	20/7/16	Updated 20/12/16
Device(s)	P162A	Wired Motion Monitor
Accessories	P145 P162C	Range of Nurse Call Leads Motion sensor keyfob
Description	The P162 detects motion of a human body by changes in heat patterns in a room. If motion is detected the alarms are transmitted by Safelink Radio or to a connected Nurse Call or Telecare system	
Directives	2014/30/EU 2014/35/EU 2011/65/EU	EMC Directive Low Voltage Permitted Materials
Classification		
Standards Applied	EN 14971:2012 EN 60950-1:2006+A2:2013 EN61000-6-3:2007+A1:2011 EN61000-6-1:2007 EN12182:2012	Risk Assessment Safety Domestic Radio Emissions Domestic Radio Immunity Assistive Technology

Declaration:

We certify that the apparatus to which this declaration relates conforms with the relevant requirements set out in the above referenced EC Directive(s) and standards

Signatory:

D Godfrey BSc(Hons) FIEE Ceng
Managing Director