

# EC Declaration of Conformity




**Manufacturer:**

iTs Designs Ltd T/A Alert-iT Care Alarms  
 Atherstone House  
 Merry Lees Industrial Estate  
 Desford  
 Leicestershire LE9 9FE  
 England

Issue Date	24/04/17	updated 24/4/17
Device(s)	P176A P176B	Bed Movement Monitor Remote Reset Button
Accessories	P168 P137 P117C P155A	Pager Pager Radio Linked Autodialler Infrastructure Interface
Description	The P176 Detects bed movement and sends a failsafe alarm by radio if this exceeds a preset value for longer than a preset time.	
Directives	<b>93/42/EEC</b> <b>2014/53/EU</b> <b>2011/65/EU</b>	Medical Device Directive RTT&E Directive RHOS2 Permitted Materials
Classification	Class 1 Medical Device	
Standards Applied	EN 14971:2012 EN 60601-1:2006++A12:2014 EN 60601-1-8:2010+A1:2015 EN62366:2015 EN12182:2012 EN 60601-1-2:2015 EN55011:2016 EN 300 220-1 V2.4.1(2012-05) ISO22196:2011	Risk Assessment Medical Device General Requirements Medical Device Alarm Systems Product Usability Engineering Technical Aids for Disabled Persons EMC ISM Radio Disturbance Radio Transmissions: Short Range Devices Anti-microbial treatment

**Declaration:**  
 We certify that the apparatus to which this declaration relates conforms with the relevant requirements set out in the above referenced EC Directive(s) and standards, for which we have sole responsibility.

**Signatory:**



D Godfrey BSc(Hons) FIEE Ceng  
 Managing Director