

EC Declaration of Conformity



Manufacturer:

iTs Designs Ltd T/A Alert-iT Care Alarms
Atherstone House,
Merry Lees Industrial Estate,
Desford,
Leicestershire, LE9 9FE,
England.

Issue Date	10/10/2016	
Device(s)	P176B	Remote Reset
Accessories	P137	Pager
Description	The P176B Remote Reset button sends an all clear signal to a paired pager. This works as part of a professional monitor system to clear an alarm call remotely.	
Directives	93/42/EEC 2014/53/EU 2011/65/EU	Medical Device Directive RTT&E Directive RHOS2 Permitted Materials
Classification	Accessory to a Class 1 Medical Device System	
Standards Applied	EN 14971:2012 EN 60601-1:2006+A112011 EN 60601-1-2:2015 EN 300 220-1 V2.4.1(2012-05) EN 60601-1-8:2010+A1:2015 EN 55011:2016 ISO22196:2011	Risk Assessment Medical Device General Requirements EMC Radio Transmissions: Short Range Devices Medical Device Alarm Systems ISM Radio Disturbance Anti-microbial treatment

Declaration:

We certify that the apparatus to which this declaration relates conforms with the relevant requirements set out in the above referenced EC Directive(s) and standards, for which we have sole responsibility.

Signatory:

A handwritten signature in black ink, appearing to read "D Godfrey".

D Godfrey BSc(Hons) FIEE Ceng
Managing Director